

CHAMELEON SOUND LEVEL METER 5 MINUTE GET UP & GO GUARANTEE

Metrosonics new Chameleon Sound Level Meter is so simple to use we guarantee you will learn how to calibrate, start and stop the collection of data, view data and clear data in 5 minutes or less - all without reading a cumbersome operator's manual.

Our guarantee is just as simple... we guarantee you can perform the 5 easy steps on the reverse side of this card, in 5 minutes or less, or you may return the Chameleon for a full refund. The unit must be returned within 30 days of delivery to the original end-user. Please contact our Customer Service Department with questions or for more details.







800.654.7778
ph.716.427.7402 • fx.716.427.8942
www.metrosonics.com



CHAMELEON SOUND LEVEL METER GET UP & GO INSTRUCTIONS


STEP 1: TURN ON Press 



STEP 2: CALIBRATE

Using cl-305 Calibrator: Connect the calibrator cable to the RS-232 connector on the back of the Chameleon (located underneath the stand). Insert the microphone into the calibrator. Press  Let the displayed reading stabilize and press 

a second time. Remove the calibrator. Using cl-304 Calibrator: Press , insert the microphone into the calibrator, turn the calibrator on, let the displayed reading stabilize and press a second time. Remove the calibrator. 

STEP 3: COLLECT DATA Press  to begin storing data. Press  a second time to stop storing data.

STEP 4: VIEW READINGS Press  repeatedly to scroll through the measurement data.

STEP 5: CLEAR MEMORY Press  CLR is displayed. Press  a second time and the memory is cleared.

CHAMELEON SOUND LEVEL METER

SETUP INSTRUCTIONS

The Chameleon Sound Level Meter can be set up for different test settings as the need arises. The setup options are as follows:

Filter: A or C-weighted

Response: Fast (F), Slow (S) or Impulse (I)

Calibration Level: 94, 102, 104 or 114 dB

*Exchange Rate: 3, 4 or 5 dB





** applies to feature packs 2, 3 & 4 only*

To select new settings, just follow the quick setup instructions on this card.

For assistance, please contact our Customer Service Department.

CHAMELEON SOUND LEVEL METER

QUICK SETUP INSTRUCTIONS

- 1: With the unit turned on, press and hold  until "SET UP" is displayed.
- 2: Press  to scroll through the available setup options and their current settings.
- 3: To change a setting, press  repeatedly until the desired setting is displayed.
- 4: Once the desired settings have been selected, press  to save the setup and return to normal operation.

NOTE: If the response rate is set to Impulse ("I"), averaging calculations will not be performed and the exchange rate settings will not be accessible during setup.